

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	AC ADAPTER
Name and address of the applicant	Shenzhen Huoniu Technology Co., Ltd. Block No.5, The 4th Industrial Zone, Xitian Community, Gongming Town, Guangming New District, Shenzhen, Guangdong, P. R. China
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	Same as applicant
Ratings and principal characteristics	Input: AC 100-240V~, 50/60 Hz, 0.35A MAX, class II Output: 4.5 - 24 Vdc, 0.01 - 2.00 A, 12 W max
Trademark (if any)	HUONIU
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	HNDI**W*
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	170117140GZU-001


This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB  
Box 1103  
SE-164 22 Kista, Sweden  
Int +46 8 750 00 00

Date: 28 March 2017



Signature:

  
Bo Berglöf



**Additional information (if necessary)**

Explanation of type designation HNDI\*\*W\*

The first symbol "\*" can be "045-240" to denote output voltage from 4.5V to 24.0V in step of 0.1 V

The second symbol "\*" can be "001-200" to denote output current from 0.01A to 2.0A in step of 0.01 A

The third symbol "\*" letter "E, U, B, X" to denote different type of integral plug part

"E" means the plug of Europe

"U" means the plug of United States of America

"B" means the plug of United Kingdom

"X" means detachable plug type

Conditions of acceptability:

1. All applied parts related requirements shall be double evaluated together with end product.
2. Temperature Testing should be performed on this component when installed in the end product.
3. Additional risk evaluation shall be conducted together with end-product.

Date: 28 March 2017

Signature: 

Intertek Testing Services Shenzhen Ltd.,  
Guangzhou Branch  
Block E, No. 7-2, Guangdong  
Software Science Park, Caipin Road,  
Guangzhou

CHINA

28 March 2017

**Handled by**  
Maarit Franzén  
**Direct telephone**  
+46 8 750 00 00  
**Reference**  
1709327  
**E-mail**  
maarit.franzen@intertek.com  
**Your reference**

**CB-certificate(s) SE-86442**

We have the pleasure to enclose the requested CB-certificate(s) and the pertaining Test Report.

Yours sincerely

Intertek Semko AB  
Product Certification

**Enclosure CB certificate(s)**